

Justification:	Function:	Electronic signature justification:	Signed: Date (yyyy-mm-dd) - Time (hh:mm):
Issued:	QA	Peter Sundsten - X-PETSUN	2020-04-21 - 10:16
Reviewed:	DD	Jon Berg - JONBER	2020-04-21 - 10:53
Approved:	DD	Fredrik Calais - FRECAL	2020-04-21 - 16:10
Released:	QA	Peter Sundsten - X-PETSUN	2020-06-16 - 09:57

This document has been electronically signed by the persons above.

Product Information

Provox® Adhesive Remover



Product description:

Provox Adhesive Remover contains a sting free solvent that helps laryngectomized patients remove Provox Adhesives and Provox Silicone Glue.

Document No: 10000038382 Edition: 04 Release date: 2020-06-16

Released

Product Information

Document ID:	PF076-02-TechInfo	Edition:	04
Manufacturer:	Atos Medical AB Kraftgatan 8, P.O. Box 183, SE-242 22 Hörby, Sweden		
Classification: (EU) 2017/745	Class I, Rule1		
Intended Use:	Provox Adhesive Remover is a single use wipe to help laryngectomized patients remove Provox Adhesives and Provox Silicone Glue.		
CE Mark:	Yes, the devices are CE marked.		
GMDN code:	60494 (Medical adhesive remover, non-sterile)		
Sterilization:	Non-sterile		
Raw material:	Cloth: Polyester and Viscous fabric Solution: Hexamethyldisiloxane, Isopropyl-myristate.		
Latex information:	Not manufactured with natural rubber latex		
Biological origin:	The device is not manufactured with materials derived from human or animal source.		
Handling and storage:	Store the product dry and away from sunlight at room temperature. Excursions permitted between 2°C - 42°C.		
Waste handling and disposal:	Waste handling and disposal should be carried out in agreement with medical practice and applicable national laws and legislations. Used product may be a potential biohazard.		
Hazardous components:	Please read the instruction for use or consult the Material Safety Data Sheet (MSDS) before handling for safe use, physical and health hazard information. The MSDS is not included with the product packaging, but can be obtained by contacting Atos Medical AB.		
Expiration date:	3 years after manufacturing.		
Packaging:	1 wipe in a bag and 50 bags in one box.		

Product Information

Devices under Basic UDI-DI: 7331791-ADH-A-000-0001-UB

REF	Name	UDI-DI
8012	Provox Adhesive Remover (50 pcs)	07331791009235
8012-18	Provox Adhesive Remover	07331791014772

Atos Medical AB compatible products:

Range	BASIC UDI-DI
Provox StabiliBase	7331791-ADH-0-000-0000-CQ
Provox Adhesive Flexiderm	7331791-ADH-0-000-0000-CQ
Provox Adhesive Optiderm	7331791-ADH-0-000-0000-CQ
Provox XtraBase Adhesive	7331791-ADH-0-000-0000-CQ
Provox Skin Barrier	7331791-ADH-A-000-0001-UB
Provox Cleaning Towel	7331791-ADH-A-000-0001-UB
Provox Silicone Glue	7331791-GEN-A-000-0003-EF
Provox Life Standard Adhesives	7331791-ADH-0-000-0001-CT
Provox Life Sensitive Adhesives	7331791-ADH-0-000-0001-CT
Provox Life Stability	7331791-ADH-0-000-0001-CT